

CERTIFICATE OF CALIBRATION

ISSUED BY AVON-DYNAMIC CALIBRATION



0199

Date of Issue 28 November 2025

Certificate Number K1048067



For:

WESTERN TOOLING SERVICES LTD.
UNIT 7, SPIBGWATER PARK
CREWS HOLE ROAD, ST GEORGE
BRISTOL
BS5 8AN

Company Number: WESTERN T

Approved Signatory
Mr S. Tregaskes

<u>Temperature:</u>	20°C ± 0.05°C	<u>Start date of calibration:</u>	25 November 2025
<u>Relative humidity:</u>	Less than 38 %	<u>Completion date of calibration:</u>	28 November 2025
<u>Description:</u>	LENGTH BAR SET		
<u>Manufacturer:</u>	PITTER TOOL & GAUGE	<u>Number of items:</u>	10
<u>Specification:</u>	BS5317:1976	<u>Serial number:</u>	1997
<u>Material:</u>	Steel	<u>Grade:</u>	1

Length measurement: Length bars up to 100 mm

The axial length of each length bar was determined by measurement by comparison with laboratory reference standards of a similar material with the length bar mounted vertically. Measurements were taken towards the edge of the measuring faces.


Length bars over 100 mm

The axial length of each length bar was determined by measuring four measurement points made by using a SIP MUL 1000 fitted with a twin probe measurement system, digital readout, and a Hewlett Packard laser interferometer. All measurement were made by holding the length bar horizontally at the Airy points.

Parallelism: The reported variation of each length bar is the difference between the maximum and minimum measurements of length obtained as above.

Flatness: The reported flatness of each length bar has been determined by using an optical flat and a monochromatic light source.

Measurement Positions: Measurement position 1 is in line with the Nominal size engraving points 2, 3 and 4 are equally spaced clockwise from position 1.



Flatness: Face A is determined from left hand side of bar from the view of the engraved nominal size.

Calibrated by: OBEES

Page 1 of 2

The results stated on this certificate relate only to the equipment calibrated.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Issued by AVON-DYNAMIC CALIBRATION

Date of Issue 28 November 2025

Certificate Number K1048067

Serial Number	Nominal Value	Length Upper Tolerance	Length Lower Tolerance	Measurement Positions				Mean Length	Uncertainty of Measurement
				Point 1	Point 2	Point 3	Point 4		
	(mm)	(μ m)	(μ m)	(μ m)	(μ m)	(μ m)	(μ m)	(mm)	(mm)
M.4819	600	3.65	-1.55	1.58	1.28	1.18	1.40	600.00136	± 0.00080
M.4959	400	2.50	-1.10	2.46	2.54 *	2.81 *	2.78 *	400.00265	± 0.00060
M.4741	300	1.40	-0.60	1.71 *	1.73 *	1.80 *	1.81 *	300.00176	± 0.00050
M.5092	200	1.40	-0.60	1.36	1.22	1.38	1.30	200.00132	± 0.00040
M.4621	100	0.85	-0.35	0.43	0.33	0.26	0.25	100.00032	± 0.00018
M.4539	80	0.70	-0.30	0.75 *	0.62	0.63	0.64	80.00066	± 0.00018
M.4686	60	0.60	-0.20	-0.51 *	-0.58 *	-0.56 *	-0.59 *	59.99944	± 0.00018
M.4781	40	0.40	-0.20	0.80 *	0.74 *	0.80 *	0.90 *	40.00081	± 0.00018
M.4347	20	0.40	-0.20	-0.12	-0.22 *	-0.14	-0.24 *	19.99982	± 0.00018
3759	10	0.40	-0.20	0.13	0.13	0.20	0.18	10.00016	± 0.00018

Serial Number	Nominal Value	Parallelism Tolerance	Results	Flatness Tolerance	Face	
					(A)	(B)
	(mm)	(μ m)	(μ m)	(μ m)	(μ m)	(μ m)
M.4819	600	0.4	0.4	0.2	0.09	0.12
M.4959	400	0.35	0.35	0.2	0.06	0.06
M.4741	300	0.25	0.1	0.2	0.12	0.12
M.5092	200	0.25	0.16	0.2	0.12	0.12
M.4621	100	0.2	0.18	0.18	0.17	0.18
M.4539	80	0.2	0.13	0.18	0.09	0.09
M.4686	60	0.18	0.08	0.15	0.12	0.15
M.4781	40	0.18	0.16	0.15	0.15	0.15
M.4347	20	0.16	0.12	0.15	0.15	0.3 *
3759	10	0.16	0.07	0.15	0.15	0.15

Uncertainty of measurement:

Less than and including 100mm - $\pm 0.18\mu$ m

Over 100mm - ± 0.20 (1 X L in m) μ m

Flatness: $\pm 0.17\mu$ m

Parallelism: $\pm 0.30\mu$ m

Calibrated using laboratory masters:

Optical Flat	ADC3951
SIP MUL 1000 incorporating a twin probe & laser interferometer	ADC3534
Master Length standards	ADC2383 ADC3529 ADC3530
Thermometer	ADC3519
Gauge block comparator	ADC3537 ADC2867

Comment: * Exceeds Tolerance

Compliance Statement:

Conformity / Non-Conformity statements are based on simple acceptance rule (ILAC-G8:09/2019) where, Acceptance Limit (AL) equals Tolerance Limit (TL). Provided that the Tolerance Uncertainty Ratio (TUR) 1:1.

-End of report-

Calibrated by: OBEES

Page 2 of 2

The results stated on this certificate relate only to the equipment calibrated.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.