

CERTIFICATE OF CALIBRATION

ISSUED BY AVON-DYNAMIC CALIBRATION

DATE OF ISSUE 13 November 2025

CERTIFICATE NUMBER K1048066



0199



CALIBRATE MEASURE INNOVATE

Page 1 of 2 Pages

Approved Signatory:
S. Tregaskes

For: Western Tooling Services Ltd
Bristol
BS5 8AN

<u>Temperature:</u>	20°C ± 2°C	<u>Date of Calibration:</u>	13 November 2025
<u>Serial No:</u>	11847	<u>Size:</u>	400 x 250 x 40mm
<u>Manufacturer:</u>	UNKNOWN		
<u>Grade:</u>	AA	<u>Material:</u>	Granite
<u>Description:</u>	A Solid Form Tri - Square.		
<u>Report:</u>	The above item has been examined and calibrated using a Master Square. A Squareness Tester incorporating a Dial Test Indicator, Electronic Comparator, Gauge Blocks and Length Bars to the tolerances specified in BS939:2007 and the results are detailed below.		

<u>Features Checked</u>	<u>Tolerance</u>	<u>Results</u>
Flatness of Face A	0.003 5mm	0.0032mm
Flatness of Face B	0.003 5mm	0.0030mm
Flatness of Side Face 1	0.008mm	0.0045mm
Flatness of Side Face 2	0.008mm	0.0048mm

CP228

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with prior written approval of the issuing laboratory.

<u>Features Checked</u>	<u>Tolerance</u>	<u>Results</u>
Squareness of Face A to Datum Face B	0.005mm	0.0040mm
Squareness of Face B to Datum Face A	0.005mm	0.0050mm
Squareness of Face A to Face 1	0.005mm	0.0020mm
Squareness of Face B to Face 1	0.005mm	0.0020mm
Squareness of Face A to Face 2	0.005mm	0.0040mm
Squareness of Face B to Face 2	0.005mm	0.0040mm

Uncertainty of Measurement: $\pm 0.005\text{mm}$

Decision Rule: Conformity / Non-Conformity statements are based on simple acceptance rule (ILAC-G8:09/2019) where, Acceptance Limit (AL) equals Tolerance Limit (TL). Provided that the Tolerance Uncertainty Ratio (TUR) $\geq 1:1$.

Acceptance Key: "*" - Indicates that the result(s) exceed the tolerance limits of the stated specification. The absence of an "*", indicates the result(s) fall within the tolerance limits of the stated specification

Calibrated using Laboratory Masters: ADC3718, ADC3319, ADC3306, ADC2890 & DCS5001

-End of Report-

CP228

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with prior written approval of the issuing laboratory.